

Reference : CHANGE NOTICE No. MB/09/03

Subject : Change in mdi Closed Sterility Test Device Stericheck (All types)

Scope : This change notification will affect the following parts/components of mdi Stericheck:

- Vent Rubber Plug - Improved grip

Background:

In our efforts to provide the user better products with more convenience and better performance, certain changes have been made in the parts/components of **mdi Stericheck**.


Need for Change:

As above.

How Does It Affect the User:

From a practical point of view the changes done will improve the product performance characteristics and ease-of-use under actual conditions of use. All performance specifications and product attributes remain unchanged.

- **Vent Rubber Plug:** New 'Butterfly' wings have been provided to the plug for easier grip and fitment.



Effective Date:

mdi Closed Sterility Test Device **Stericheck** (all types) manufactured after October 24, 2009 will have said changes.

In case you have any queries, please feel free to contact our Technical Support Team at 'mdi@vsnl.com'.

Quality Assurance
Advanced Microdevices Pvt. Ltd.