

Reference : CHANGE NOTICE No. MB/09/04

Subject : Change in mdi Closed Sterility Test Device Stericheck Type PC (All membrane types)

Scope : This change notification will affect the following parts/components of mdi Stericheck:

- Needle of Stericheck-PC - Improved

Background:

In our efforts to provide the user better products with more convenience and better performance, certain changes have been made in the needles of mdi Stericheck Type PC.


Need for Change:

As above.

How Does It Affect the User:

From a practical point of view the changes done will improve the product performance characteristics and ease-of-use under actual conditions of use. All performance specifications and product attributes remain unchanged.

- Needle of Stericheck-PC: The needle will now have only side opening.



Effective Date:

mdi Closed Sterility Test Device Stericheck Type PC manufactured after October 26, 2009 will have said changes.

In case you have any queries, please feel free to contact our Technical Support Team at 'mdi@vsnl.com'.

Quality Assurance
Advanced Microdevices Pvt. Ltd.