

ADVANCED MICRODEVICES PVT. LTD.

20 - 21, INDUSTRIAL AREA, AMBALA CANTT - 133 001, INDIA

Phone: + 91 - 171 - 2699290, 2699471, 2699274

Website: www.mdimembrane.com

Change Notice No.	Notification date	Implementation date
CN/F/24/009	August 16, 2024	August 24, 2024

Subject : Updation of filtration area of mdi Harvester Funnels

Scope : This change will affect the products with following catalog numbers: (# can be any

alphabet or numeral):

FHXG5238#######

FHXG5238########

Background:

This advance change notification is initiated to provide the user updated information regarding the products they are using. The filtration area of **mdi** *Harvester* Funnels with catalog numbers as mentioned in the scope has been updated in the Certificate of Quality (CoQ).

Need for Change:

mdi 500mL and 1000mL *Harvester* Funnels have a nominal filtration area of 400cm² and 1200cm² respectively and the same is reflecting in respective product datasheet.

However, a typographical error was identified in Certificate of Quality (CoQ) of **mdi** 500mL and 1000mL *Harvester* Funnels. The filtration area is wrongly mentioned as 200cm² (for the 500mL *Harvester* Funnel) and 500cm² (for the 1000mL *Harvester* Funnel) in Certificate of Quality (CoQ).

Hence, to rectify this error, the Certificate of Quality (CoQ) shall be corrected.

No changes have been made to the materials of construction or performance parameters or manufacturing process or critical dimensions or traceability.

How Does It Affect the User:

From a practical point of view this change specifically corrects the typographical error related to the filtration area mentioned in the Certificate of Quality (CoQ). All product specifications including materials of construction as well as performance specifications shall remain unchanged. The accompanying Certificate of Quality will reflect the said change.

Implementation of Change:

mdi Harvester Funnels with catalog numbers as mentioned in the scope, released after the implementation date will have the mentioned changes.

In case you have any queries, please feel free to contact our Technical Support Team at 'info@mdimembrane.com'.

m

Head - Quality Assurance
Advanced Microdevices Pvt. Ltd.