

# Certificate of Quality

The Polyethersulfone Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

## **AseptiPrime KSPES Membrane Capsule Filters**

Catalog No. : DKX75136AAXB101  
 Type : DK  
 Pore Size : 0.1 µm (0.5 µm + 0.1 µm)  
 Lot Number : DK9694H SI.No. 056  
 Manufacturing Date : 2024 - 08  
 Expiry Date : 2029 - 08

### SPECIFICATION

|                                      |                                                   |
|--------------------------------------|---------------------------------------------------|
| <b>Length</b>                        | 1"                                                |
| <b>Filter Media</b>                  | Polyethersulfone Membrane                         |
| <b>Drainage Layers</b>               | Polyester                                         |
| <b>Housing</b>                       | Polypropylene                                     |
| <b>Differential Pressure</b>         | < 4Kg/cm <sup>2</sup> at 30 °C                    |
| <b>Maximum Operating Temperature</b> | 80 °C @ < 2 Kg/cm <sup>2</sup>                    |
| <b>Sterilization</b>                 | 3 autoclaving cycles at 125 °C of 30 minutes each |

### LOT RELEASE CRITERIA

#### 100% Integrity Tested

: The Capsule filter has been tested for integrity by Air Diffusion Flow Test and Bubble Point Test using purified water.

Diffusion flows were: ≤ 1.8 ml/min @ 3.52 kg/cm<sup>2</sup>

Bubble point was: ≥ 65 psi (4.48 Bar).

#### Typical Water Flow Rate

: 300 ml/min @ 0.14 Kg/cm<sup>2</sup> @ 27 °C

#### Microbial Challenge Test

: Retains ≥ 10<sup>7</sup> organisms/cm<sup>2</sup> of *Acholeplasma laidlawii* (ATCC 23206).

### VALIDATED FOR

#### Bubble point (50% IPA)

: The filter is certified/validated for integrity by Bubble point test using 50% IPA/Water solution. Bubble point ≥ 26 psi (1.79 Bar)

#### Bacterial Endotoxin

: Aqueous extracts exhibit < 0.25 EU/mL as established by Limulus Amebocyte Lysate (LAL) test as per USP <85>.

#### Heat Stability

: Maintains integrity after 3 autoclaving cycles at 125 °C of 30 minutes each.

#### Extractable

: Within limits as specified in USP.

#### Oxidizable matter

: Passes test as per USP.

#### Biosafety

: Passes Biological Reactivity Tests, *In Vivo* for Class VI plastic as described in USP <88>.

#### Cytotoxicity

: Passes Biological Reactivity Tests, *In Vitro* as described in USP <87>.

#### Indirect Food Additives

: Passes as per FDA 21CFR 177.1520(a)1(i).

#### Particle Release

: Passes test as per USP <788>, "Particulate matter in Injections".

#### Fiber Release

: Complies with FDA 21CFR 210.3(b)(6).

#### Total Organic Carbon

: Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified volume of water for injection.

#### Conductivity

: Meets USP <645> limit of 1.3 µS/cm at 25 °C for water conductivity after flushing specified volume of water for injection.

### CUSTOMER SUPPORT

**mdi** offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/CAP/001-04



Head of Quality Assurance

Issue Date: 27-Aug-24

### **Advanced Microdevices Pvt. Ltd.**

Jawahargarh Road,  
 Village-Tepla, Ambala, INDIA.  
 Tel: +91-171-2699290/2699274  
 Website: www.mdimembrane.com  
 Email: info@mdimembrane.com

**An ISO 9001 Company**