

The Hydrophilic PVDF Membrane filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

## **AseptiCap WS- $\gamma$ Hydrophilic PVDF Membrane Filters**

|                    |   |
|--------------------|---|
| Catalog No.        | : VWSX1001BBRX102   |
| Type               | : VWS   |
| Pore Size          | : 0.2 $\mu\text{m}$ (0.45 $\mu\text{m}$ + 0.2 $\mu\text{m}$ ) |
| Lot Number         | : IV5884L   |
| Manufacturing Date | : 2024 - 12   |
| Expiry Date        | : 2029 - 12   |

### **SPECIFICATION**

|                                     |  |
|-------------------------------------|--|
| Membrane                            | : Hydrophilic PVDF                                   |
| Housing                             | : Polypropylene                                      |
| Filter Diameter                     | : 50 mm  |
| Effective Filtration Area (Nominal) | : 20 $\text{cm}^2$                                   |
| Burst pressure                      | : > 8 $\text{Kg/cm}^2$                               |
| Sterilization                       | : Can be sterilized by Gamma Irradiation upto 50 kGy |

### **LOT RELEASE CRITERIA**

The above lot meets the following lot release criteria:

|                                 |   |
|---------------------------------|---|
| <b>100% Integrity Tested</b>    | : The filter has been tested for integrity by Bubble Point test using DI water. Bubble point value with DI water was: $\geq 50$ psi (3.44 Bar)          |
| <b>Typical Water Flow Rate</b>  | : 150 ml/min @ 0.70 $\text{kg/cm}^2$ @ 27 $^{\circ}\text{C}$  |
| <b>Microbial Challenge Test</b> | : Retains $\geq 10^7$ organisms/ $\text{cm}^2$ of <i>B. diminuta</i> ATCC 19146 challenge as per ASTM F838 methodology.                                 |
| <b>VALIDATED FOR</b>            |   |
| <b>Bubble point (50% IPA)</b>   | : The filter is certified/validated for integrity by Bubble point test using 50% IPA/Water solution. Bubble point $\geq 18$ psi (1.24 Bar)              |
| <b>Bacterial Endotoxin</b>      | : Filtrate meets the USP requirements for Sterile WFI of $\leq 0.25$ EU/ml as determined by Limulus Amebocyte Lysate (LAL) test.                        |
| <b>Biosafety</b>                | : Passes Biological Reactivity Tests, <i>In Vivo</i> for Class VI plastic as described in USP <88>.   |
| <b>Cytotoxicity</b>             | : Passes Biological Reactivity Tests, <i>In Vitro</i> as described in USP <87>.   |
| <b>Indirect Food Additives</b>  | : Passes as per FDA 21CFR 177.1520(a)1(i).  |
| <b>Particle Release</b>         | : Passes test as per USP <788>, "Particulate matter in Injections".   |
| <b>Fiber Release</b>            | : Complies with FDA 21CFR 210.3(b)(6).  |
| <b>Total Organic Carbon</b>     | : Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified volume of water for injection.                                     |
| <b>Conductivity</b>             | : Meets USP <645> limit of 1.3 $\mu\text{S/cm}$ at 25 $^{\circ}\text{C}$ for water conductivity after flushing specified volume of water for injection. |

### **PRECAUTIONS**

1. During handling, avoid contamination of outlet.
2. If pressure required to maintain the required flow becomes too high, the filter unit should be changed.

### **CUSTOMER SUPPORT**

**mdi** offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/ILF/016-03



Head of Quality Assurance

Issue Date: 16-Dec-24

### **Advanced Microdevices Pvt. Ltd.**

Jawahargarh Road,  
Village-Tepla, Ambala, INDIA.  
Tel: +91-171-2699290/2699274  
Website: [www.mdimembrane.com](http://www.mdimembrane.com)  
Email: [info@mdimembrane.com](mailto:info@mdimembrane.com)

**An ISO 9001 Company**