

Certificate of Quality

The Polyethersulfone Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

AseptiCap KL-γ PES Membrane Capsule Filters

Catalog No. : DKLX5702DDXX301
 Type : DKL-S
 Pore Size : 0.45 μm
 Lot Number : DK9304K SI.No. 234
 Ster. No. : R2564
 Date of Sterilization : 2024 - 11
 Expiry Date : 2026 - 11

SPECIFICATION

| | |
|--------------------------------------|---|
| Length | 8" |
| Filter Media | Polyethersulfone Membrane |
| Drainage Layers | Polyester |
| Housing | Polypropylene |
| Differential Pressure | < 4Kg/cm ² at 30 °C |
| Maximum Operating Temperature | 80 °C @ < 2Kg/cm ² |
| Sterilization | Pre sterilized by Gamma Irradiation dose between 25 kGy to 40 kGy |

LOT RELEASE CRITERIA

100% Integrity Tested

: The capsule filter has been tested for integrity by Bubble point test using DI water. Bubble point value with DI water was: ≥ 30 psi (2.07 Bar)

Typical Water Flow Rate

: 19 lpm @ 0.70 Kg/cm² @ 27 °C

VALIDATED FOR

Bubble point (50% IPA)

: The filter is certified/validated for integrity by Bubble point test using 50% IPA/Water solution. Bubble point ≥ 10 psi (0.69 Bar)

Sterility

: The sterilization process has been validated to assure Sterility Assurance Level (SAL) of 10⁻⁶ in accordance with ISO 11137.

Microbial Retention

: Retains microbial challenge of *S. marcescens* (ATCC 14756).

Bacterial Endotoxin

: Aqueous extracts exhibit < 0.25 EU/mL as established by Limulus Amebocyte Lysate (LAL) test as per USP <85>.

Extractable

: Within limits as specified in USP.

Oxidizable matter

: Passes test as per USP.

Biosafety

: Passes Biological Reactivity Tests, *In Vivo* for Class VI plastic as described in USP <88>.

Cytotoxicity

: Passes Biological Reactivity Tests, *In Vitro* as described in USP <87>.

Indirect Food Additives

: Passes as per FDA 21CFR 177.1520(a)1(i).

Particle Release

: Passes test as per USP <788>, "Particulate matter in Injections".

Fiber Release

: Complies with FDA 21CFR 210.3(b)(6).

Total Organic Carbon

: Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified volume of water for injection.

Conductivity

: Meets USP <645> limit of 1.3 μS/cm at 25 °C for water conductivity after flushing specified volume of water for injection.

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/CAP/004-05



Head of Quality Assurance

Issue Date: 07-Nov-24

Advanced Microdevices Pvt. Ltd.

Jawahargarh Road,
 Village-Tepla, Ambala, INDIA.
 Tel: +91-171-2699290/2699274
 Website: www.mdimembrane.com
 Email: info@mdimembrane.com

An ISO 9001 Company