

Certificate of Quality

The Nylon-66 Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using *validated production processes*.

AseptiCap NNZ Nylon-66 Membrane Capsule Filters

Catalog No. : DNNZ5301BBXX101

Type : DNNZ
Pore Size : 0.2 µm

Lot Number : DN9934L SI.No. 025

Manufacturing Date : 2024 - 12 Expiry Date : 2029 - 12

SPECIFICATION

Length	5"
Filter Media	Positively charged Nylon-66 Membrane
Drainage Layers	Polyester
Housing	Polypropylene
Differential Pressure	< 4Kg/cm² at 30 °C
Maximum Operating Temperature	80 °C @ < 2 Kg/cm²

LOT RELEASE CRITERIA

100% Integrity Tested

: The capsule filter has been tested for integrity by Bubble point test using DI water.

Bubble point value with DI water was: ≥ 50 psi (3.44 Bar)

Typical Water Flow Rate

: 5 lpm @ 0.70 Kg/cm² @ 27 °C

Microbial Challenge Test

: Retains $\geq 10^7$ organisms/cm² of *B. diminuta* ATCC 19146 challenge as per ASTM

F838 methodology.

VALIDATED FOR

Bubble point (50% IPA)

: The filter is certified/validated for integrity by Bubble point test using 50% IPA/Water solution. Bubble point ≥ 17 psi (1.17 Bar)

Sterilization

: Autoclavable @ 125 °C for 30 minutes.

Bacterial Endotoxin

: Aqueous extracts exhibit < 0.25 EU/mL as established by Limulus Amebocyte

Lysate (LAL) test as per USP <85>. Within limits as specified in USP.

Extractable Biosafety

: Passes Biological Reactivity Tests, In Vivo for Class VI plastic as described in USP

<88>

Indirect Food Additives

: Passes as per FDA 21CFR 177.1520(a)1(i).

Particle Release

: Passes test as per USP <788>, "Particulate matter in Injections".

Fiber Release

: Complies with FDA 21CFR 210.3(b)(6).

Total Organic Carbon

: Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified

volume of water for injection.

Conductivity

: Meets USP <645> limit of 1.3 µS/cm at 25 °C for water conductivity after flushing specified volume of water for injection.

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/CAP/009-02

Om

Head of Quality Assurance

Issue Date: 17-Dec-24

Advanced Microdevices Pvt. Ltd.

Jawahargarh Road,

Village-Tepla, Ambala, INDIA.
Tel: +91-171-2699290/2699274
Website: www.mdimembrane.com

Website: www.mdimembrane.com Email: info@mdimembrane.com

An ISO 9001 Company