

Certificate of Quality

The Polyethersulfone Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

AseptiCap KSO PES Membrane Capsule Filters

Catalog No. : DKOX5201EEXX101
Type : DKSO
Pore Size : 0.2 μm (0.45 μm + 0.2 μm)
Lot Number : DK7772I SI.No. 007

SPECIFICATION

| | |
|--------------------------------------|--|
| Length | 2" |
| Filter Media | Polyethersulfone Membrane |
| Drainage Layers | Polypropylene |
| Differential Pressure | < 4Kg/cm ² at 30 °C |
| Housing | Polypropylene |
| Maximum Operating Temperature | 80 °C @ < 2 Kg/cm ² |
| Sterilization | 25 autoclaving cycles at 125 °C of 30 minutes each |

LOT RELEASE CRITERIA

- 100% Integrity Tested** : The capsule filter has been tested for integrity by Bubble point test using DI water. Bubble point value with DI water was: \geq 50 psi (3.44 Bar)
The capsule filter is also certified for integrity by Bubble point test using 50% IPA/Water solution.
Bubble point with 50% IPA/Water solution is \geq 16 psi (1.10 Bar)
- Water Flow Rate** : \geq 3.5 lpm @ 0.70 Kg/cm² @ 27 °C
- Microbial Challenge Test** : Retains $\geq 10^7$ organisms/cm² of *B. diminuta* ATCC 19146 challenge as per ASTM F838-05 methodology.

VALIDATED FOR

- Heat Stability** : Maintains integrity after 25 autoclaving cycles at 125 °C of 30 minutes each.
- Extractable** : Within limits as specified in USP.
- Oxidizable matter** : Passes test as per USP.
- Biosafety** : Passes Biological Reactivity Tests, *In Vivo* for Class VI plastic as described in USP <88>.
- Indirect Food Additives** : Passes as per FDA 21CFR 177.1520(a)1(i).
- Particle Release** : Passes test as per USP <788>, "Particulate matter in Injections".
- Fiber Release** : Complies with FDA 21CFR 210.3(b)(6).

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.



Head of Quality Assurance

Issue Date: 11-Sep-17

Advanced Microdevices Pvt. Ltd.

21, Industrial Area, Ambala Cantt, INDIA,
Tel: +91-171-2699290/ 2699274
Website: www.mdimembrane.com
Email: info@mdimembrane.com

An ISO 9001 Company