

## Certificate of Quality

The Polyethersulfone Membrane Cartridge Filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

## AseptiSure KS PES Membrane Cartridge Filters

Catalog No. : CPKX5301G0XX101

Type : CPPKS

Pore Size : 0.2 μm (0.45 μm + 0.2 μm)
Lot Number : CK5547I SI.No. 012

## **SPECIFICATION**

Length	5"
Filter Media	Polyethersulfone Membrane
Drainage Layers	Polyester
Plastic Components	Polypropylene
Differential Pressure	< 3.5 Kg/cm² at 25°C
Maximum operating Temperature	80 °C at < 2 Kg/cm²
Reverse Pressure	< 0.7 Kg/cm² at 25°C
Sterilization	25 Autoclave/ Steam sterilization cycles at 121°C for 30 minutes each

## **LOT RELEASE CRITERIA**

**100% Integrity Tested** : The cartridge filter has been tested for integrity by Air Diffusion Flow test and Bubble

Point test using DI water.

Diffusion flows with DI water were: ≤ 15 ml/min @ 2.60 kg/cm² Bubble point value with DI water was: ≥ 50 psi (3.44 Bars)

The cartridge filter is also certified for integrity by Bubble point test using 50%

IPA/Water solution.

Bubble point with 50% IPA/Water solution is ≥ 16 psi (1.10 Bars)

**Water Flow Rate** : ≥ 7.0 lpm @ 0.70 Kg/cm² @ 27 °C

Microbial Challenge Test : Retains ≥ 10<sup>7</sup> organisms/cm² of *B. diminuta* ATCC 19146 challenge as per

ASTM F838-05 methodology.

**VALIDATED FOR** 

**Heat Stability** : Maintains integrity after 25 autoclave/ Steam sterilization cycles at 121°C for 30

minutes each.

**Extractable**: Within limits as specified in USP.

Oxidizable matter : Passes test as per USP.

**Bacterial Endotoxins** : Filtrate meets the USP requirements for Sterile WFI of ≤ 0.25 EU/ml as determined

by Limulus Amebocyte Lysate (LAL) test.

Biosafety : Passes Biological Reactivity Tests, In Vivo for Class VI plastic as described

in USP <88>.

**Cytotoxicity**: Passes Biological Reactivity Tests, *In Vitro* as described in USP <87>.

**Indirect Food Additives**: Passes as per FDA 21CFR 177.1520(a)1(i).

Particle Release : Passes test as per USP <788>, "Particulate matter in Injections".

Fiber Release : Complies with FDA 21CFR 210.3(b)(6).

**CUSTOMER SUPPORT** 

**mdi** offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

Com

Head of Quality Assurance

Issue Date: 11-Sep-17

Advanced Microdevices Pvt. Ltd.

21, Industrial Area, Ambala Cantt, INDIA, Tel: +91-171-2699290/ 2699274
Website: www.mdimembrane.com
Email: info@mdimembrane.com

An ISO 9001 Company