

Certificate of Quality

The RBC Separation Membrane has been manufactured in a **mdi** facility in compliance with ISO-9001 regulations using validated production processes.

RBC SEPARATION MEMBRANE

TYPE : FR-2 (0.7)

LOT NO. : R-107073

PHYSICAL TESTS

Thickness (μm) : 688
Tare weight (mg/cm^2) : 11.1
Size (mm) : 150x300

FUNCTIONAL TESTS

Apply 40 μl whole blood.
Run with 120 μl PBS 0.01M, pH 7.4
(5x30 mm FR2 on 15 μ CNPC)

Hemolysis : No
Visible leakage of RBC on
NC membrane : No

Ashu

Head of Quality Control

Advanced Microdevices Pvt. Ltd.

21, Industrial Area, Ambala Cantt, INDIA,
Tel: +91-171-2699290/ 2699274/ 2699471
Fax: +91-171-2699221/ 2699008
Website: www.mdimembrane.com
Email: mdi@vsnl.com

