

Certificate of Quality

The Polyethersulfone Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

AseptiCap Nano PES Membrane Capsule Filters

Catalog No. : DKNX53S7DDXX101

Type : DKN

Pore Size : Sterilizing Grade

Lot Number : DK6694L SI.No. 095

Manufacturing Date : 2024 - 12 Expiry Date : 2029 - 12

SPECIFICATION

Length	5"
Filter Media	Polyethersulfone Membrane
Drainage Layers	Polyester
Housing	Polypropylene
Differential Pressure	< 4Kg/cm² at 30 °C
Maximum Operating Temperature	80 °C @ < 2 Kg/cm²
Sterilization	3 autoclaving cycles at 125 °C of 30 minutes each

LOT RELEASE CRITERIA 100% Integrity Tested

: The capsule filter has been tested for integrity by Air Diffusion Flow test using DI water. Diffusion flow with DI water was: ≤ 2.5 ml/min @ 1.41 kg/cm² (20 psi)

Typical Water Flow Rate Microbial Challenge Test

: 8.0 lpm @ 0.70 Kg/cm² @ 27 °C : LRV > 7 for *B. diminuta* ATCC 19146 challenge as per ASTM F838-05

methodology.

VALIDATED FOR

Bacterial Endotoxin : Aqueous extracts exhibit < 0.25 EU/mL as established by Limulus Amebocyte

Lysate (LAL) test as per USP <85>.

Heat Stability : Maintains integrity after 3 autoclaving cycles at 125 °C of 30 minutes each.

Extractable : Within limits as specified in USP.

Oxidizable matter : Passes test as per USP.

Blosafety : Passes Biological Reactivity Tests, *In Vivo* for Class VI plastic as described in USP

<88>.

Indirect Food Additives : Passes as per FDA 21CFR 177.1520(a)1(i).

Particle Release : Passes test as per USP <788>, "Particulate matter in Injections".

Fiber Release : Complies with FDA 21CFR 210.3(b)(6).

Total Organic Carbon: Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified

volume of water for injection.

Conductivity : Meets USP <645> limit of 1.3 μ S/cm at 25 °C for water conductivity after flushing

specified volume of water for injection.

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/CAP/001-04

Om

Head of Quality Assurance

Issue Date: 17-Dec-24

Advanced Microdevices Pvt. Ltd.

Jawahargarh Road, Village-Tepla, Ambal

Village-Tepla, Ambala, INDIA.
Tel: +91-171-2699290/2699274
Website: www.mdimembrane.com
Email: info@mdimembrane.com

An ISO 9001 Company