

The Hydrophilic PVDF Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using *validated production processes*.

Certificate of Quality

<i>AseptiCap WL</i> Hy Filters	drophilic PV	DF Membrane Capsule	
Catalog No.	: DWLX5201D	DXX101	
Туре	: DWL		
Pore Size	: 0.2 µm		
Lot Number	: DV5334L	SI.No. 069	
Manufacturing Date	: 2024 - 12		
Expiry Date	: 2029 - 12		

SPECIFICATION

SPECIFICATION					
	Length		2"		
	Filter Media		Hydrophilic PVDF Membrane		
	Drainage Layers		Polyester		
	Housing		Polypropylene		
	Differential Pressure		< 4Kg/cm ² at 30 °C		
	Maximum Operating Temperature		80 °C @ < 2 Kg/cm ²		
	Sterilization		2 Autoclaving cycles at 125 °C of 30 minutes each		
LOT R	ELEASE CRITERIA				
100%	Integrity Tested	: The capsule filter has been tested for integrity by Bubble point test using DI water. Bubble point value with DI water was: ≥ 50 psi (3.44 Bar)			
Typica	al Water Flow Rate	: 4.0 lpm @ 0.70 Kg/cm² @ 27 °C			
Micro	bial Challenge Test	Retains $\geq 10^7$ organisms/cm ² of <i>B. diminuta</i> ATCC 19146 challenge as per ASTM F838 methodology.			
VALID	ATED FOR				
Bubbl	e point (50% IPA)	: The filter is certified/validated for integrity by Bubble point test using 50% IPA/Water solution. Bubble point ≥ 18 psi (1.24 Bar)			
Heat S	stability	: Maintains integrity after 2 autoclaving cycles at 125 °C of 30 minutes each.			
Extrac	table	: Within limits as specified in USP.			
Oxidiz	able matter	Passes test as per USP.			
Bacte	rial Endotoxin	 Filtrate meets the USP requirements for Sterile WFI of ≤ 0.25 EU/ml as determined by Limulus Amebocyte Lysate (LAL) test. 			
Biosa	fety	: Passes Biological Reactivity Tests, <i>In Vivo</i> for Class VI plastic as described in USP <88>.			
Cytoto	oxicity	: Passes E	: Passes Biological Reactivity Tests, In Vitro as described in USP <87>.		
Indire	ct Food Additives	Passes as per FDA 21CFR 177.1520(a)1(i).			
Partic	le Release	: Passes test as per USP <788>, "Particulate matter in Injections".			
Fiber	Release	Complies with FDA 21CFR 210.3(b)(6).			
Total	Organic Carbon		: Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified volume of water for injection.		
Condu	ıctivity		Meets USP <645> limit of 1.3 μ S/cm at 25 °C for water conductivity after flushing specified volume of water for injection.		

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/CAP/012-03

han

Head of Quality Assurance Issue Date: 16-Dec-24 Advanced Microdevices Pvt. Ltd.

Jawahargarh Road, Village-Tepla, Ambala, INDIA. Tel: +91-171-2699290/2699274 Website: www.mdimembrane.com Email: info@mdimembrane.com

An ISO 9001 Company