

Certificate of Quality

The Polypropylene syringe filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations.

Polypropylene Syringe Filters

Catalog No. : SYPP0307MNXX104
Type : SY13PP
Pore Size : 5 µm
Lot Number : SP9907G

SPECIFICATION

Membrane : Polypropylene
Housing : Polypropylene
Filter Diameter : 13 mm
Filtration Area : 0.8 cm²
Burst pressure : > 14 Kg/cm²

QUALITY ASSURANCE LOT RELEASE CRITERIA

Each lot has been sampled, tested and released by the Quality Assurance Department for the following characteristics:

- Integrity
- Flow Rate Performance
- Burst Pressure
- Leakage Test

BIOSAFETY

: Passes biological test for plastics Class VI as per USP and complies with Code of Federal Regulations FDA 21CFR 177.1520(a)1(i).

EXPIRY

: Non degradable polymeric materials of construction with no expiry when stored at room temperature in a dry place.

PRECAUTIONS

1. During handling, avoid contamination of outlet.
2. If pressure required to maintain the required flow becomes too high, the filter unit should be changed.

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.



Head of Quality Assurance

Issue Date: 24-Jul-17

Advanced Microdevices Pvt. Ltd.
21, Industrial Area, Ambala Cantt, INDIA,
Tel: +91-171-2699290/ 2699274
Website: www.mdimembrane.com
Email: info@mdimembrane.com

An ISO 9001 Company