

Certificate of Quality

The Polyethersulfone Membrane Capsule filters have been manufactured in a **mdi** facility in compliance with **ISO 9001** regulations using **validated production processes**.

BioPro KSO- PES Membrane Capsule Filters

Catalog No. : DBKO5301DDXX301

Type : DBKSO-S Pore Size : 0.2 µm

Lot Number : DK9334L SI.No. 056

Ster. No. : R6624
Date of Sterilization : 2024 - 12
Expiry Date : 2026 - 12

SPECIFICATION

| Length | 5" |
|-------------------------------|---|
| Filter Media | Polyethersulfone Membrane |
| Drainage Layers | Polypropylene |
| Housing | Polypropylene |
| Differential Pressure | < 4Kg/cm² at 30 °C |
| Maximum Operating Temperature | 80 °C @ < 2Kg/cm ² |
| Sterilization | Pre sterilized by Gamma Irradiation dose between 25 kGy to 40 kGy |

LOT RELEASE CRITERIA

100% Integrity Tested : The capsule filter has been tested for integrity by Bubble point test using DI water.

Bubble point was: \geq 30 psi (2.07 Bar).

Typical Water Flow Rate : 9.0 lpm @ 0.70 Kg/cm² @ 27 °C

VALIDATED FOR

Sterility: The sterilization process has been validated to assure Sterility Assurance Level

(SAL) of 10⁻⁶ in accordance with ISO 11137.

Microbial Challenge Test : Retains $\geq 10^5$ organisms/cm² of *B. diminuta* ATCC 19146 challenge.

Bacterial Endotoxin : Aqueous extracts exhibit < 0.25 EU/mL as established by Limulus Amebocyte

Lysate (LAL) test as per USP <85>.

Extractable : Within limits as specified in USP.

Oxidizable matter : Passes test as per USP.

Biosafety: Passes Biological Reactivity Tests, In Vivo for Class VI plastic as described in USP

<88>.

Cytotoxicity : Passes Biological Reactivity Tests, *In Vitro* as described in USP <87>.

indirect Food Additives : Passes as per FDA 21CFR 177.1520(a)1(i).

Particle Release : Passes test as per USP <788>, "Particulate matter in Injections".

Fiber Release : Complies with FDA 21CFR 210.3(b)(6).

Total Organic Carbon : Meets USP <643> limit of 500 ppb for total organic carbon after flushing specified

volume of water for injection.

Conductivity : Meets USP <645> limit of 1.3 µS/cm at 25 °C for water conductivity after flushing

specified volume of water for injection.

CUSTOMER SUPPORT

mdi offers its unique interdisciplinary skills to provide solutions to specific problems. Please contact our factory or the local application specialist.

T. No.: COQ/CAP/004-05

Com

Head of Quality Assurance Issue Date: 16-Dec-24 Advanced Microdevices Pvt. Ltd.

Jawahargarh Road,

Email:

Village-Tepla, Ambala, INDIA.
Tel: +91-171-2699290/2699274
Website: www.mdimembrane.com

info@mdimembrane.com

- ICO 0001 Comment

An ISO 9001 Company